

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2200

Name: RITTER 1  
Sterilizer ID: V2178198  
08/04/2022 11:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.0 F 0.0  
2:00 171.4 F 3.6  
4:00 182.4 F 7.6  
6:00 202.7 F 12.3  
8:00 217.4 F 6.2  
10:00 233.0 F 11.4  
12:00 232.0 F 7.8  
14:00 244.3 F 12.8  
16:00 243.3 F 12.0  
18:00 255.7 F 18.5  
20:00 266.7 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.4  
2:00 272.2 F 28.9  
3:00 271.3 F 28.5  
4:00 272.3 F 29.4  
  
Min 271.3 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 227.7 F 0.0  
10:00 227.9 F 0.0  
15:00 229.8 F 0.0  
20:00 230.9 F 0.0  
25:00 229.5 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 21:35  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:51  
CYCLE COMPLETE

08/04/2022 11:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_