
Midmark M11 - v1.0.5
Total Cycles: 3713

Name: RITTER2
Sterilizer ID: V2313228
06/18/2024 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.9 F 0.0
2:00 100.9 F 0.4
4:00 116.8 F 2.3
6:00 144.0 F 5.1
8:00 166.9 F 8.6
10:00 186.9 F 12.4
12:00 206.0 F 5.2
14:00 218.0 F 9.3
16:00 229.5 F 11.5
18:00 235.0 F 10.6
20:00 234.9 F 8.4
22:00 248.1 F 14.8
24:00 258.8 F 20.8
26:00 268.7 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.9
1:00 271.7 F 30.0
2:00 271.6 F 30.1
3:00 271.5 F 30.0
4:00 271.5 F 29.9

Min 271.5 F 29.8
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 226.8 F 0.0
10:00 191.8 F 0.0
15:00 181.8 F 0.0
20:00 181.3 F 0.0
25:00 178.7 F 0.0
30:00 182.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 27:51
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:06:31
CYCLE COMPLETE

06/18/2024 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____