
Midmark M11 - v1.0.5
Total Cycles: 54

Name: RITTER2
Sterilizer ID: V2313228
05/31/2021 06:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 73.9 F 0.2
4:00 102.4 F 0.9
6:00 141.9 F 3.1
8:00 151.8 F 6.2
10:00 168.7 F 9.9
12:00 188.8 F 8.0
14:00 209.2 F 5.1
16:00 220.5 F 8.8
18:00 230.8 F 12.8
20:00 233.1 F 9.1
22:00 241.0 F 11.2
24:00 244.5 F 12.8
26:00 256.1 F 19.0
28:00 267.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.0 F 28.9
2:00 272.2 F 29.0
3:00 272.4 F 29.2
4:00 272.4 F 29.4

Min 271.4 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.2 F 0.0
5:00 188.7 F 0.0
10:00 189.0 F 0.0
15:00 198.3 F 0.0
20:00 200.2 F 0.0
25:00 197.7 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:20
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:07:14
CYCLE COMPLETE

05/31/2021 07:34 AM
ACCEPTED: X
REJECTED:
Operator: U01