
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004649
9:23 03/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:30
PRESSURIZING 1:32
STERILIZING 2:36
136.4 C 220kPa 2:36
136.3 C 223kPa 3:06
136.4 C 222kPa 3:36
136.4 C 223kPa 4:06
136.4 C 223kPa 4:36
136.3 C 222kPa 5:06
136.4 C 223kPa 5:36
136.3 C 222kPa 6:06
136.3 C 222kPa 6:06
VENTING 6:06
AIR DRYING 6:53
CYCLE COMPLETE 7:02

Digital Signature #
AAF5D72D79BBE36E
