
Midmark M11 - v1.0.5
Total Cycles: 4048

Name: RITTER2
Sterilizer ID: V2313228
10/23/2024 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.4 F 0.0
2:00 111.4 F 0.0
4:00 118.3 F 0.0
6:00 149.4 F 1.4
8:00 169.7 F 4.7
10:00 188.1 F 8.2
12:00 205.0 F 12.0
14:00 218.4 F 6.4
16:00 225.8 F 10.8
18:00 231.0 F 7.8
20:00 242.8 F 12.8
22:00 240.5 F 11.2
24:00 252.1 F 17.0
26:00 262.2 F 22.9
28:00 271.9 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 271.8 F 29.9
2:00 272.3 F 30.0
3:00 272.6 F 29.9
4:00 272.8 F 29.8

Min 271.4 F 29.5
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 236.1 F 0.0
10:00 206.7 F 0.0
15:00 190.4 F 0.0
20:00 172.9 F 0.0
25:00 171.0 F 0.0
30:00 170.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:11
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:47
CYCLE COMPLETE

10/23/2024 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____