
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001731
9:51 15/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:50
STERILIZING 3:51
136.0 C 220kPa 3:51
136.4 C 223kPa 4:21
136.1 C 221kPa 4:51
136.4 C 224kPa 5:21
136.4 C 221kPa 5:51
136.3 C 224kPa 6:21
136.4 C 224kPa 6:51
136.2 C 221kPa 7:21
136.2 C 221kPa 7:21
VENTING 7:21
AIR DRYING 8:11
CYCLE COMPLETE 23:32

Digital Signature #
FBCD22D64E8D32A9
