
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000319
10:59 04/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:49
STERILIZING 3:59
136.0 C 228kPa 3:59
136.2 C 232kPa 4:29
136.4 C 233kPa 4:59
136.4 C 233kPa 5:29
136.1 C 230kPa 5:59
136.4 C 233kPa 6:29
136.4 C 233kPa 6:59
136.3 C 233kPa 7:29
136.3 C 233kPa 7:29
VENTING 7:29
AIR DRYING 8:24
CYCLE COMPLETE 23:24

Digital Signature #
2F0F76D433B91B3D
