
Midmark M11 - v1.0.0
Total Cycles: 4041

Name: RITTER 1
Sterilizer ID: V2178198
11/20/2023 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 168.4 F 2.8
4:00 176.5 F 6.4
6:00 212.3 F 10.7
8:00 220.2 F 4.1
10:00 228.9 F 8.9
12:00 239.7 F 12.4
14:00 238.6 F 10.5
16:00 238.4 F 8.2
18:00 249.4 F 14.8
20:00 260.6 F 21.1
22:00 270.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 29.0
2:00 271.9 F 28.4
3:00 271.6 F 28.6
4:00 272.8 F 28.9

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 217.8 F 0.0
10:00 222.0 F 0.0
15:00 228.0 F 0.0
20:00 227.4 F 0.0
25:00 224.5 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:48
CYCLE COMPLETE

11/20/2023 01:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____