

-----  
Midmark M11 - v1.0.5  
Total Cycles: 813

Name: RITTER2  
Sterilizer ID: V2313228  
02/17/2022 09:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.2 F 0.0  
2:00 128.3 F 1.2  
4:00 135.7 F 3.7  
6:00 158.2 F 7.5  
8:00 183.0 F 12.0  
10:00 210.2 F 5.3  
12:00 224.1 F 10.0  
14:00 230.6 F 8.7  
16:00 239.3 F 12.4  
18:00 239.3 F 11.0  
20:00 252.6 F 17.5  
22:00 263.7 F 24.2  
24:00 271.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.5  
1:00 272.4 F 30.0  
2:00 271.3 F 29.4  
3:00 272.4 F 30.1  
4:00 271.9 F 29.5

Min 271.3 F 29.3  
Max 272.4 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 204.9 F 0.0  
10:00 223.0 F 0.0  
15:00 218.9 F 0.0  
20:00 216.8 F 0.0  
25:00 217.4 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 24:04  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:02:24  
CYCLE COMPLETE

02/17/2022 10:37 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02