
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002379
14:04 13/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:01
STERILIZING 3:00
136.0 C 220kPa 3:00
136.3 C 225kPa 3:30
136.4 C 226kPa 4:00
136.4 C 225kPa 4:30
136.4 C 225kPa 5:00
136.4 C 225kPa 5:30
136.1 C 221kPa 6:00
136.1 C 222kPa 6:30
136.1 C 222kPa 6:30
VENTING 6:30
AIR DRYING 7:20
CYCLE COMPLETE 37:20

Digital Signature #
842BB1E4FA567D2A
