
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000641
11:26 09/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:17
STERILIZING 4:23
136.0 C 228kPa 4:23
136.1 C 230kPa 4:53
136.4 C 233kPa 5:23
136.3 C 233kPa 5:53
136.4 C 233kPa 6:23
136.2 C 230kPa 6:53
136.4 C 230kPa 7:23
136.3 C 232kPa 7:53
136.3 C 232kPa 7:53
VENTING 7:53
AIR DRYING 8:49
CYCLE COMPLETE 23:49

Digital Signature #
9E9EE9B83A851CE2
