

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002353  
14:05 04/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:12  
PRESSURIZING 2:18  
STERILIZING 3:21  
136.0 C 220kPa 3:21  
136.4 C 224kPa 3:51  
136.0 C 222kPa 4:21  
136.2 C 223kPa 4:51  
136.0 C 221kPa 5:21  
136.4 C 222kPa 5:51  
136.2 C 223kPa 6:21  
136.2 C 222kPa 6:51  
136.2 C 222kPa 6:51  
VENTING 6:51  
AIR DRYING 7:42  
CYCLE COMPLETE 37:42  
-----

Digital Signature #  
185F6AD3F2C91F82  
-----