

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2229

Name: RITTER2  
Sterilizer ID: V2313228  
04/11/2023 10:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.2 F 0.0  
2:00 171.5 F 3.3  
4:00 184.8 F 7.1  
6:00 212.1 F 12.0  
8:00 223.1 F 6.7  
10:00 235.0 F 11.8  
12:00 233.2 F 9.2  
14:00 237.4 F 9.5  
16:00 246.9 F 14.3  
18:00 258.8 F 20.9  
20:00 270.4 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.5 F 29.8  
2:00 271.0 F 29.1  
3:00 272.4 F 30.0  
4:00 271.7 F 29.4  
  
Min 271.0 F 29.1  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.9 F 0.0  
5:00 229.8 F 0.0  
10:00 217.5 F 0.0  
15:00 226.6 F 0.0  
20:00 225.8 F 0.0  
25:00 228.5 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 20:49  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:10  
CYCLE COMPLETE

04/11/2023 11:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_