
Midmark M11 - v1.0.0
Total Cycles: 1168

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2021 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.1 F 0.0
2:00 142.6 F 1.1
4:00 161.4 F 3.5
6:00 182.6 F 6.8
8:00 202.5 F 11.1
10:00 214.3 F 4.1
12:00 228.3 F 8.9
14:00 238.2 F 12.7
16:00 239.0 F 10.6
18:00 235.7 F 8.3
20:00 249.1 F 14.8
22:00 260.5 F 21.2
24:00 270.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.1
2:00 272.1 F 28.6
3:00 271.5 F 28.4
4:00 272.7 F 29.2

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 202.5 F 0.0
10:00 195.1 F 0.0
15:00 194.2 F 0.0
20:00 192.8 F 0.0
25:00 207.0 F 0.0
Dry Cycle Aborted

FILLING: 1:06
HEATING: 24:49
STERILIZING: 4:00
VENTING: 3:22
DRYING: 29:24
TOTAL CYCLE: 01:02:41

09/14/2021 10:38 AM
ACCEPTED: X
REJECTED:
Operator: U01