
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004813
10:43 06/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:00
STERILIZING 4:08
136.4 C 224kPa 4:08
136.1 C 222kPa 4:38
136.1 C 220kPa 5:08
136.4 C 224kPa 5:38
136.3 C 223kPa 6:08
136.4 C 225kPa 6:38
136.3 C 222kPa 7:08
136.1 C 221kPa 7:38
136.1 C 221kPa 7:38
VENTING 7:38
AIR DRYING 8:38
CYCLE COMPLETE 9:06

Digital Signature #
6C55612360CC913A
