
Midmark M11 - v1.0.5
Total Cycles: 2283

Name: RITTER2
Sterilizer ID: V2313228
04/21/2023 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 174.6 F 3.4
4:00 190.3 F 7.6
6:00 209.9 F 12.4
8:00 222.0 F 6.9
10:00 234.3 F 12.1
12:00 236.3 F 9.7
14:00 237.4 F 9.3
16:00 249.3 F 15.3
18:00 261.3 F 22.0
20:00 272.0 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.1 F 29.2
2:00 271.8 F 29.5
3:00 271.9 F 29.1
4:00 271.8 F 29.5

Min 271.2 F 28.8
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 211.3 F 0.0
10:00 212.5 F 0.0
15:00 217.5 F 0.0
20:00 227.9 F 0.0
25:00 228.5 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

04/21/2023 12:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____