

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1995

Name: RITTER2  
Sterilizer ID: V2313228  
02/01/2023 02:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
2:00 191.0 F 4.5  
4:00 205.8 F 10.3  
6:00 220.4 F 5.6  
8:00 235.2 F 11.9  
10:00 237.2 F 10.0  
12:00 236.2 F 9.0  
14:00 251.5 F 16.5  
16:00 264.1 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.2 F 29.4  
2:00 271.3 F 28.6  
3:00 272.2 F 29.6  
4:00 271.8 F 29.1  
  
Min 271.1 F 28.6  
Max 272.3 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
5:00 225.4 F 0.0  
10:00 224.8 F 0.0  
15:00 226.0 F 0.0  
20:00 228.6 F 0.0  
25:00 230.9 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 2:11  
HEATING: 17:55  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:20  
CYCLE COMPLETE

02/01/2023 03:52 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_