

-----  
Midmark M11 - v1.0.0  
Total Cycles: 978

Name: RITTER 1  
Sterilizer ID: V2178198  
07/07/2021 03:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 214.5 F 0.0  
2:00 203.4 F 4.0  
4:00 222.2 F 9.1  
6:00 229.9 F 8.4  
8:00 236.7 F 9.9  
10:00 235.4 F 8.4  
12:00 246.4 F 13.4  
14:00 244.6 F 12.3  
16:00 257.4 F 19.1  
18:00 268.8 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.0  
1:00 272.3 F 28.5  
2:00 271.9 F 28.6  
3:00 271.5 F 28.0  
4:00 272.3 F 28.6  
  
Min 271.2 F 27.9  
Max 272.4 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.5 F 0.0  
5:00 221.2 F 0.0  
10:00 224.7 F 0.0  
15:00 222.1 F 0.0  
20:00 212.2 F 0.0  
25:00 212.5 F 0.0  
30:00 208.0 F 0.0  
  
Drying Complete

FILLING: 1:54  
HEATING: 18:58  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:06  
CYCLE COMPLETE

07/07/2021 04:10 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01