

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 001436  
16:59 14/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:45  
STERILIZING 4:09  
136.4 C 226kPa 4:09  
136.3 C 225kPa 4:39  
136.4 C 226kPa 5:09  
136.4 C 224kPa 5:39  
136.3 C 226kPa 6:09  
136.2 C 224kPa 6:39  
136.2 C 224kPa 7:09  
136.4 C 226kPa 7:39  
136.4 C 226kPa 7:39  
VENTING 7:39  
AIR DRYING 8:34  
CYCLE COMPLETE 11:21  
-----

Digital Signature #  
E2D9A0EA7D6E1C5A  
-----