
Midmark M11 - v1.0.5
Total Cycles: 4339

Name: RITTER2
Sterilizer ID: V2313228
02/06/2025 04:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.3 F 0.0
2:00 95.8 F 0.6
4:00 113.2 F 2.4
6:00 138.3 F 5.3
8:00 165.3 F 9.1
10:00 189.3 F 13.1
12:00 209.6 F 5.0
14:00 220.7 F 8.9
16:00 232.1 F 13.2
18:00 233.8 F 9.8
20:00 236.8 F 9.6
22:00 245.2 F 13.8
24:00 256.7 F 20.0
26:00 267.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 272.7 F 30.1
2:00 271.4 F 29.8
3:00 272.5 F 30.0
4:00 271.3 F 29.8

Min 271.2 F 29.5
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 225.0 F 0.0
10:00 227.4 F 0.0
15:00 200.9 F 0.0
20:00 191.8 F 0.0
25:00 204.2 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 27:28
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:06:08
CYCLE COMPLETE

02/06/2025 05:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____