

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000815  
17:02 02/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:45  
STERILIZING 4:12  
136.4 C 227kPa 4:12  
136.2 C 230kPa 4:42  
136.3 C 230kPa 5:12  
136.1 C 229kPa 5:42  
136.4 C 229kPa 6:12  
136.1 C 230kPa 6:42  
136.0 C 230kPa 7:12  
136.2 C 231kPa 7:42  
136.2 C 231kPa 7:42  
VENTING 7:42  
AIR DRYING 8:31  
CYCLE COMPLETE 17:57  
-----

Digital Signature #  
5B64F4FBBB1AD7DE  
-----