
Midmark M11 - v1.0.0
Total Cycles: 1603

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2022 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.4 F 0.0
2:00 137.0 F 2.1
4:00 157.9 F 5.4
6:00 186.6 F 9.6
8:00 210.6 F 4.6
10:00 220.1 F 7.8
12:00 232.7 F 12.6
14:00 235.0 F 9.2
16:00 238.8 F 9.8
18:00 246.2 F 13.3
20:00 258.0 F 19.7
22:00 268.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.9
2:00 271.4 F 28.4
3:00 272.4 F 28.9
4:00 271.5 F 28.2

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 231.8 F 0.0
10:00 220.6 F 0.0
15:00 217.9 F 0.0
20:00 217.8 F 0.0
25:00 225.7 F 0.0
30:00 233.5 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:07
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

02/07/2022 05:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____