
Midmark M11 - v1.0.5
Total Cycles: 4128

Name: RITTER2
Sterilizer ID: V2313228
11/15/2024 04:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.4 F 0.0
2:00 155.0 F 1.3
4:00 172.5 F 4.8
6:00 198.9 F 9.0
8:00 212.8 F 11.9
10:00 224.6 F 8.2
12:00 235.1 F 13.1
14:00 236.5 F 10.5
16:00 235.8 F 8.8
18:00 249.2 F 15.5
20:00 260.7 F 22.0
22:00 271.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.2 F 29.5
2:00 272.6 F 29.9
3:00 271.7 F 29.6
4:00 272.3 F 29.6

Min 271.6 F 29.2
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 223.8 F 0.0
10:00 228.4 F 0.0
15:00 204.1 F 0.0
20:00 195.5 F 0.0
25:00 192.5 F 0.0
30:00 195.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:39
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

11/15/2024 05:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____