
Midmark M11 - v1.0.0
Total Cycles: 2402

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2022 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.5 F 0.0
2:00 116.9 F 0.4
4:00 138.6 F 1.9
6:00 174.6 F 4.3
8:00 207.2 F 8.3
10:00 223.3 F 13.2
12:00 224.8 F 6.7
14:00 233.8 F 11.2
16:00 231.8 F 7.6
18:00 243.9 F 13.3
20:00 242.4 F 11.5
22:00 254.3 F 17.8
24:00 265.4 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.3
2:00 271.6 F 28.5
3:00 272.5 F 29.4
4:00 272.1 F 28.9

Min 271.3 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 231.1 F 0.0
10:00 247.5 F 0.0
15:00 237.4 F 0.0
20:00 225.8 F 0.0
25:00 226.4 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 25:51
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

10/05/2022 12:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____