
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001213
9:46 16/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:51
STERILIZING 5:12
136.4 C 224kPa 5:12
136.3 C 224kPa 5:42
136.4 C 224kPa 6:12
136.3 C 224kPa 6:42
136.5 C 224kPa 7:12
136.4 C 224kPa 7:42
136.4 C 220kPa 8:12
136.0 C 221kPa 8:42
136.0 C 221kPa 8:42
VENTING 8:42
AIR DRYING 9:36
CYCLE COMPLETE 24:36

Digital Signature #
1112A7ECE64E4B19
