
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000581
12:40 12/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:19
STERILIZING 3:25
136.0 C 227kPa 3:25
136.3 C 232kPa 3:55
136.2 C 228kPa 4:25
136.1 C 231kPa 4:55
136.2 C 231kPa 5:25
136.4 C 233kPa 5:55
136.3 C 233kPa 6:25
136.3 C 232kPa 6:55
136.3 C 232kPa 6:55
VENTING 6:55
AIR DRYING 7:45
CYCLE COMPLETE 17:42

Digital Signature #
B9653F985A736B1A
