

-----  
Midmark M11 - v1.0.5  
Total Cycles: 914

Name: RITTER2  
Sterilizer ID: V2313228  
03/28/2022 11:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.7 F 0.0  
2:00 164.5 F 0.3  
4:00 193.4 F 3.5  
6:00 209.2 F 8.8  
8:00 233.2 F 9.8  
10:00 234.0 F 11.6  
12:00 236.5 F 10.0  
14:00 236.3 F 9.0  
16:00 251.2 F 16.5  
18:00 263.6 F 23.8  
20:00 271.5 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.7 F 29.9  
2:00 271.5 F 29.2  
3:00 272.3 F 30.1  
4:00 272.3 F 29.7  
  
Min 271.2 F 29.2  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.2 F 0.0  
5:00 225.6 F 0.0  
10:00 228.0 F 0.0  
15:00 228.9 F 0.0  
20:00 231.6 F 0.0  
25:00 225.6 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:21  
CYCLE COMPLETE

03/28/2022 12:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_