
Midmark M11 - v1.0.0
Total Cycles: 2253

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2022 05:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.6 F 0.0
2:00 194.8 F 0.7
4:00 209.7 F 4.6
6:00 225.9 F 9.7
8:00 231.9 F 8.2
10:00 241.8 F 12.4
12:00 241.1 F 11.0
14:00 239.8 F 10.2
16:00 254.7 F 17.5
18:00 266.9 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.1 F 28.3
2:00 272.6 F 28.8
3:00 271.5 F 27.9
4:00 272.6 F 28.6

Min 271.5 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 228.6 F 0.0
10:00 241.0 F 0.0
15:00 238.0 F 0.0
20:00 225.3 F 0.0
25:00 220.9 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

08/25/2022 06:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____