
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003396
9:32 06/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:56
STERILIZING 2:52
136.5 C 220kPa 2:52
136.4 C 224kPa 3:22
136.4 C 222kPa 3:52
136.3 C 221kPa 4:22
136.4 C 220kPa 4:52
136.4 C 223kPa 5:22
136.3 C 223kPa 5:52
136.4 C 224kPa 6:22
136.4 C 224kPa 6:22
VENTING 6:22
AIR DRYING 7:13
CYCLE COMPLETE 20:10

Digital Signature #
422367AE7F7158EA
