
Midmark M11 - v1.0.0
Total Cycles: 3667

Name: RITTER 1
Sterilizer ID: V2178198
08/17/2023 07:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 75.0 F 0.0
4:00 83.1 F 0.0
6:00 113.2 F 0.6
8:00 142.2 F 3.3
10:00 167.5 F 6.4
12:00 195.2 F 9.9
14:00 213.2 F 10.7
16:00 221.3 F 5.7
18:00 230.9 F 9.4
20:00 241.7 F 13.3
22:00 236.4 F 9.5
24:00 239.6 F 9.8
26:00 245.0 F 12.7
28:00 256.5 F 18.7
30:00 266.7 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.5 F 28.2
2:00 271.3 F 28.2
3:00 271.5 F 28.2
4:00 271.2 F 28.1

Min 271.2 F 28.1
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 239.7 F 0.0
10:00 227.3 F 0.0
15:00 229.3 F 0.0
20:00 228.9 F 0.0
25:00 228.4 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:29
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

08/17/2023 08:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____