
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 003638
11:01 10/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:04
STERILIZING 4:04
136.4 C 223kPa 4:04
136.2 C 224kPa 4:34
136.2 C 222kPa 5:04
136.4 C 225kPa 5:34
136.3 C 225kPa 6:04
136.3 C 224kPa 6:34
136.1 C 222kPa 7:04
136.3 C 223kPa 7:34
136.3 C 223kPa 7:34
VENTING 7:34
AIR DRYING 8:27
CYCLE COMPLETE 9:15

Digital Signature #
FFB28EFB58204AFB
