
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002484
15:49 06/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:59
STERILIZING 2:59
136.4 C 225kPa 2:59
136.2 C 223kPa 3:29
136.4 C 226kPa 3:59
136.2 C 223kPa 4:29
136.3 C 224kPa 4:59
136.4 C 226kPa 5:29
136.4 C 223kPa 5:59
136.2 C 225kPa 6:29
136.2 C 225kPa 6:29
VENTING 6:29
AIR DRYING 7:24
CYCLE COMPLETE 23:26

Digital Signature #
663D38B6B8580133
