
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000691
7:43 16/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:25
STERILIZING 3:33
136.4 C 232kPa 3:33
136.4 C 232kPa 4:03
136.4 C 232kPa 4:33
136.5 C 232kPa 5:03
136.2 C 231kPa 5:33
136.2 C 229kPa 6:03
136.0 C 231kPa 6:33
136.2 C 231kPa 7:03
136.2 C 231kPa 7:03
VENTING 7:03
AIR DRYING 7:50
CYCLE COMPLETE 8:18

Digital Signature #
E34F224DF3D5F702
