

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 002331  
11:01 29/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:08  
STERILIZING 3:09  
136.4 C 220kPa 3:09  
136.0 C 223kPa 3:39  
135.9 C 222kPa 4:09  
136.4 C 225kPa 4:39  
136.4 C 225kPa 5:09  
136.4 C 224kPa 5:39  
136.4 C 222kPa 6:09  
136.1 C 221kPa 6:39  
136.1 C 221kPa 6:39  
VENTING 6:39  
AIR DRYING 7:31  
CYCLE COMPLETE 13:50  
-----

Digital Signature #  
7AEC40A98E51E599  
-----