
Midmark M11 - v1.0.0
Total Cycles: 1552

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2022 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.3 F 0.0
2:00 184.2 F 2.6
4:00 186.4 F 7.0
6:00 210.1 F 12.6
8:00 223.0 F 6.6
10:00 235.1 F 11.7
12:00 234.9 F 8.7
14:00 240.5 F 10.3
16:00 246.9 F 13.7
18:00 259.5 F 20.6
20:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.1
2:00 271.8 F 28.4
3:00 271.8 F 28.7
4:00 272.9 F 29.1

Min 271.3 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 217.9 F 0.0
10:00 226.8 F 0.0
15:00 229.6 F 0.0
20:00 229.2 F 0.0
25:00 229.5 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

01/24/2022 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____