
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005740
11:31 12/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:05
STERILIZING 3:01
136.0 C 220kPa 3:01
136.4 C 223kPa 3:31
136.2 C 222kPa 4:01
136.4 C 225kPa 4:31
136.2 C 223kPa 5:01
136.2 C 223kPa 5:31
136.3 C 224kPa 6:01
136.4 C 225kPa 6:31
136.4 C 225kPa 6:31
VENTING 6:31
AIR DRYING 7:19
CYCLE COMPLETE 8:05

Digital Signature #
CAAE7F0BCA3CE22F
