
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000747
7:07 27/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 2:01
STERILIZING 2:53
136.0 C 227kPa 2:53
136.4 C 231kPa 3:23
136.1 C 230kPa 3:53
136.1 C 230kPa 4:23
136.4 C 230kPa 4:53
136.2 C 229kPa 5:23
136.4 C 233kPa 5:53
136.4 C 232kPa 6:23
136.4 C 232kPa 6:23
VENTING 6:23
AIR DRYING 7:10
CYCLE COMPLETE 7:16

Digital Signature #
3ECDD80B033C931F
