
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003975
14:20 08/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:38
STERILIZING 4:53
136.5 C 221kPa 4:53
136.1 C 222kPa 5:23
136.3 C 224kPa 5:53
136.4 C 224kPa 6:23
136.2 C 223kPa 6:53
136.3 C 224kPa 7:23
136.4 C 224kPa 7:53
136.4 C 224kPa 8:23
136.4 C 224kPa 8:23
VENTING 8:23
AIR DRYING 9:18
CYCLE COMPLETE 9:19

Digital Signature #
32626B812AB0718E
