
Midmark M11 - v1.0.0
Total Cycles: 4112

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2023 11:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.8 F 0.0
2:00 123.3 F 1.1
4:00 127.0 F 3.0
6:00 149.1 F 5.8
8:00 193.2 F 9.0
10:00 205.4 F 13.1
12:00 213.5 F 5.3
14:00 225.5 F 9.3
16:00 235.4 F 13.5
18:00 236.2 F 9.9
20:00 239.7 F 9.4
22:00 245.7 F 13.1
24:00 256.4 F 18.9
26:00 266.5 F 25.0
28:00 272.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.6 F 29.1
2:00 272.1 F 29.2
3:00 272.2 F 28.9
4:00 272.4 F 29.4

Min 271.6 F 28.7
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.1 F 0.0
5:00 221.0 F 0.0
10:00 230.5 F 0.0
15:00 229.3 F 0.0
20:00 225.4 F 0.0
25:00 224.3 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 28:29
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:17
CYCLE COMPLETE

12/13/2023 12:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____