
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001661
15:01 11/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:33
STERILIZING 3:28
136.0 C 220kPa 3:28
136.2 C 224kPa 3:58
136.0 C 223kPa 4:28
136.2 C 225kPa 4:58
136.2 C 223kPa 5:28
136.4 C 224kPa 5:58
136.4 C 226kPa 6:28
136.4 C 225kPa 6:58
136.4 C 225kPa 6:58
VENTING 6:59
AIR DRYING 7:49
CYCLE COMPLETE 11:15

Digital Signature #
B7235D025325F320
