
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002181
11:19 29/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:21
STERILIZING 3:15
136.0 C 221kPa 3:15
136.3 C 225kPa 3:45
136.4 C 225kPa 4:15
136.4 C 223kPa 4:45
136.4 C 224kPa 5:15
136.3 C 226kPa 5:45
136.0 C 223kPa 6:15
136.2 C 222kPa 6:45
136.2 C 222kPa 6:45
VENTING 6:45
AIR DRYING 7:36
CYCLE COMPLETE 23:20

Digital Signature #
65DE0394DA57B090
