
Midmark M11 - v1.0.5
Total Cycles: 4032

Name: RITTER2
Sterilizer ID: V2313228
10/16/2024 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.2 F 0.0
2:00 134.0 F 0.8
4:00 145.1 F 3.4
6:00 174.2 F 7.2
8:00 198.1 F 11.7
10:00 215.2 F 5.0
12:00 223.2 F 10.0
14:00 232.5 F 8.7
16:00 240.4 F 12.2
18:00 238.9 F 10.8
20:00 251.7 F 17.0
22:00 262.9 F 23.5
24:00 272.8 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.8
1:00 272.6 F 29.8
2:00 272.8 F 30.2
3:00 271.8 F 30.1
4:00 272.1 F 29.8

Min 271.5 F 29.7
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 205.3 F 0.0
10:00 226.2 F 0.0
15:00 242.4 F 0.0
20:00 229.5 F 0.0
25:00 226.3 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:60
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:03:31
CYCLE COMPLETE

10/16/2024 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____