
Midmark M11 - v1.0.5
Total Cycles: 1963

Name: RITTER2
Sterilizer ID: V2313228
01/24/2023 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.2 F 0.0
2:00 166.9 F 0.2
4:00 178.5 F 1.0
6:00 200.0 F 5.5
8:00 217.8 F 10.5
10:00 225.5 F 6.6
12:00 238.4 F 12.3
14:00 238.6 F 10.5
16:00 237.3 F 9.5
18:00 251.9 F 16.7
20:00 264.1 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.5
2:00 271.6 F 28.8
3:00 272.5 F 29.7
4:00 272.1 F 29.2

Min 271.5 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 224.8 F 0.0
10:00 226.9 F 0.0
15:00 229.8 F 0.0
20:00 226.4 F 0.0
25:00 227.4 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

01/24/2023 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____