
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000004
10:33 19/12/2019
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:39
PRESSURIZING 0:47
STERILIZING 1:48
136.4 C 220kPa 1:48
136.4 C 222kPa 2:18
136.4 C 223kPa 2:48
136.4 C 223kPa 3:18
136.3 C 223kPa 3:48
136.4 C 224kPa 4:18
136.3 C 222kPa 4:48
136.3 C 221kPa 5:18
136.3 C 221kPa 5:18
VENTING 5:18
AIR DRYING 6:05
CYCLE COMPLETE 6:19

Digital Signature #
FD5ED8B6FBA6777C
