
Midmark M11 - v1.0.5
Total Cycles: 2831

Name: RITTER2
Sterilizer ID: V2313228
09/05/2023 01:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.1 F 0.0
2:00 130.2 F 0.0
4:00 153.2 F 2.1
6:00 183.4 F 5.6
8:00 197.5 F 9.6
10:00 219.0 F 9.0
12:00 226.3 F 8.7
14:00 237.1 F 13.5
16:00 237.7 F 10.8
18:00 236.2 F 9.2
20:00 249.6 F 15.6
22:00 260.4 F 21.9
24:00 270.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 271.6 F 29.6
2:00 272.1 F 29.9
3:00 272.1 F 30.0
4:00 272.0 F 30.0

Min 271.0 F 29.0
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 209.1 F 0.0
10:00 210.1 F 0.0
15:00 209.8 F 0.0
20:00 208.2 F 0.0
25:00 211.0 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:45
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:03:18
CYCLE COMPLETE

09/05/2023 02:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____