
Midmark M11 - v1.0.5
Total Cycles: 673

Name: RITTER2
Sterilizer ID: V2313228
01/05/2022 12:55 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 160.7 F 1.4
4:00 175.7 F 4.7
6:00 198.4 F 9.7
8:00 220.0 F 6.5
10:00 234.2 F 11.9
12:00 236.7 F 10.2
14:00 236.1 F 9.1
16:00 251.6 F 16.6
18:00 264.3 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 271.8 F 28.8
2:00 272.1 F 29.0
3:00 272.5 F 29.4

Min 271.6 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 223.0 F 0.0
5:00 232.4 F 0.0
10:00 221.4 F 0.0
Dry Cycle Aborted

FILLING: 2:06
HEATING: 19:47
STERILIZING: 3:00
VENTING: 2:44
DRYING: 10:52
TOTAL CYCLE: 00:38:29

01/05/2022 01:33 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____