
Midmark M11 - v1.0.5
Total Cycles: 225

Name: RITTER2
Sterilizer ID: V2313228
07/20/2021 06:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.2 F 0.0
2:00 75.7 F 0.0
4:00 86.8 F 0.8
6:00 105.5 F 3.0
8:00 132.6 F 5.8
10:00 159.7 F 9.2
12:00 180.2 F 12.9
14:00 203.0 F 4.2
16:00 214.7 F 7.7
18:00 226.2 F 11.7
20:00 229.9 F 8.0
22:00 241.4 F 13.2
24:00 241.0 F 11.7
26:00 252.9 F 17.8
28:00 263.5 F 24.0
30:00 272.6 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.5
1:00 272.7 F 30.2
2:00 271.5 F 29.4
3:00 272.3 F 30.3
4:00 272.1 F 29.8

Min 271.1 F 29.3
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 202.0 F 0.0
10:00 207.4 F 0.0
15:00 200.1 F 0.0
20:00 194.2 F 0.0
25:00 188.8 F 0.0
30:00 196.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:18
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:08:20
CYCLE COMPLETE

07/20/2021 07:32 AM
ACCEPTED: X
REJECTED:
Operator: U01