
Midmark M11 - v1.0.0
Total Cycles: 3034

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2023 07:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.5 F 0.0
2:00 79.8 F 0.0
4:00 111.2 F 0.0
6:00 132.9 F 0.4
8:00 151.3 F 3.1
10:00 175.8 F 6.1
12:00 200.7 F 9.6
14:00 215.3 F 12.3
16:00 221.3 F 6.5
18:00 231.5 F 10.5
20:00 235.5 F 8.0
22:00 240.9 F 11.7
24:00 237.7 F 9.4
26:00 250.8 F 15.7
28:00 261.9 F 22.1
30:00 271.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.7
3:00 272.5 F 29.2
4:00 271.5 F 28.5

Min 271.4 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 201.5 F 0.0
10:00 208.9 F 0.0
15:00 221.5 F 0.0
20:00 241.2 F 0.0
25:00 222.4 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:33
CYCLE COMPLETE

03/27/2023 08:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____