
Midmark M11 - v1.0.0
Total Cycles: 4926

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2024 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.8 F 0.0
2:00 156.7 F 0.1
4:00 174.8 F 0.4
6:00 195.5 F 2.4
8:00 210.0 F 5.9
10:00 222.2 F 9.5
12:00 231.8 F 13.4
14:00 233.6 F 8.6
16:00 242.8 F 13.4
18:00 241.5 F 11.0
20:00 237.4 F 9.1
22:00 250.7 F 15.3
24:00 261.4 F 21.4
26:00 271.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 271.9 F 28.5
2:00 272.5 F 28.8
3:00 272.2 F 28.5
4:00 271.8 F 28.2

Min 271.2 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 222.5 F 0.0
10:00 243.5 F 0.0
15:00 230.2 F 0.0
20:00 224.6 F 0.0
25:00 226.4 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:50
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:05:31
CYCLE COMPLETE

09/20/2024 12:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____