
Midmark M11 - v1.0.5
Total Cycles: 4049

Name: RITTER2
Sterilizer ID: V2313228
10/23/2024 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 83.7 F 0.0
2:00 83.9 F 0.0
4:00 87.9 F 0.0
6:00 111.7 F 0.1
8:00 158.9 F 0.2
10:00 187.3 F 0.5
12:00 211.1 F 3.7
14:00 217.6 F 6.9
16:00 226.2 F 10.3
18:00 231.0 F 9.8
20:00 236.9 F 10.8
22:00 234.4 F 8.5
24:00 242.6 F 11.9
26:00 244.9 F 13.1
28:00 256.1 F 19.0
30:00 266.2 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.4 F 29.7
2:00 271.9 F 29.1
3:00 271.4 F 29.3
4:00 272.5 F 29.7

Min 271.1 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 202.6 F 0.0
10:00 211.5 F 0.0
15:00 203.5 F 0.0
20:00 212.5 F 0.0
25:00 226.0 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:47
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:10:29
CYCLE COMPLETE

10/23/2024 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____